

**New Online** Views **27,547** | Citations **0** | Altmetric **1587** | Comments

PDF | Share | Cite | Permissions | Comment

**Viewpoint**

**ONLINE FIRST** **FREE**

May 6, 2020

# Interpreting Diagnostic Tests for SARS-CoV-2

Nandini Sethuraman, MD<sup>1</sup>; Sundararaj Stanleyraj Jeremiah, MD<sup>2</sup>; Akihide Ryo, MD, PhD<sup>2</sup>

» Author Affiliations | Article Information

JAMA. Published online May 6, 2020. doi:10.1001/jama.2020.8259


Related Articles

Interviews

The pandemic of coronavirus disease 2019 (COVID-19) continues to affect much of the world. Knowledge of diagnostic tests for severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) is still evolving, and a clear understanding of the nature of the tests and interpretation of their findings is important. This Viewpoint describes how to interpret 2 types of diagnostic tests commonly in use for SARS-CoV-2 infections—reverse transcriptase–polymerase chain reaction (RT-PCR) and IgM and IgG enzyme-linked immunosorbent assay (ELISA)—and how the results may vary over time (**Figure**).

Download  

**Figure. Estimated Variation Over Time in Diagnostic Tests for Detection of SARS-CoV-2 Infection Relative to Symptom Onset**



Estimated time intervals and rates of viral detection are based on data from several published reports. Because of variability-

ty in values among studies, estimated time intervals should be considered approximations and the probability of detection of SARS-CoV-2 infection is presented qualitatively. SARS-CoV-2 indicates severe acute respiratory syndrome coronavirus 2; PCR, polymerase chain reaction.

<sup>a</sup>Detection only occurs if patients are followed up proactively from the time of exposure.

<sup>b</sup>More likely to register a negative than a positive result by PCR of a nasopharyngeal swab.

---

Advertisement

## Detection of Viral RNA by RT-PCR

Thus far, the most commonly used and reliable test for diagnosis of COVID-19 has been the RT-PCR test performed using nasopharyngeal swabs or other upper respiratory tract specimens, including throat swab or, more recently, saliva. A variety of RNA gene targets are used by different manufacturers, with most tests targeting 1 or more of the envelope (*env*), nucleocapsid (*N*), spike (*S*), RNA-dependent RNA polymerase (*RdRp*), and *ORF1* genes. The sensitivities of the tests to individual genes are comparable according to comparison studies except the RdRp-SARSr (Charité) primer probe, which has a slightly lower sensitivity likely due to a mismatch in the reverse primer.<sup>1</sup>

In most individuals with symptomatic COVID-19 infection, viral RNA in the nasopharyngeal swab as measured by the cycle threshold (Ct) becomes detectable as early as day 1 of symptoms and peaks within the first week of symptom onset. The Ct is the number of replication cycles required to produce a fluorescent signal, with lower Ct values representing higher viral RNA loads. A Ct value less than 40 is clinically reported as PCR positive. This positivity starts to decline by week 3 and subsequently becomes undetectable. However, the Ct values obtained in severely ill hospitalized patients are lower than the Ct values of mild cases, and PCR positivity may persist beyond 3 weeks after illness onset when most mild cases will yield a negative result.<sup>2</sup> However, a "positive" PCR result reflects only the detection of viral RNA and does not necessarily indicate presence of viable virus.<sup>3</sup>

In some cases, viral RNA has been detected by RT-PCR even beyond week 6 following the first positive test. A few cases have also been reported positive after 2 consecutive negative PCR tests performed 24 hours apart. It is unclear if this is a testing error, reinfection, or reactivation. In a study of 9 patients, attempts to isolate the virus in culture were not successful beyond day 8 of illness onset, which correlates with the decline of infectivity beyond the first week.<sup>3</sup> That is in part why the "symptom-based strategy" of the Centers for Disease Control and Prevention (CDC) indicates that health care workers can return to work, if "at least 3 days (72 hours) have passed since recovery defined as resolution of fever without the use of fever-reducing medications and improvement in respiratory symptoms (e.g., cough, shortness of breath); and, at least 10 days have passed since symptoms first appeared."<sup>4</sup>

The timeline of PCR positivity is different in specimens other than nasopharyngeal swab. PCR positivity declines more slowly in sputum and may still be positive after nasopharyngeal swabs are negative.<sup>3</sup> In one study, PCR positivity in stool was observed in 55 of 96 (57%) infected patients and remained positive in stool beyond nasopharyngeal swab by a median of 4 to 11 days, but was unrelated to clinical severity.<sup>2</sup> Persistence of PCR in sputum and stool was found to be similar as assessed by Wölfel et al.<sup>3</sup>

In a study of 205 patients with confirmed COVID-19 infection, RT-PCR positivity was highest in bronchoalveolar lavage specimens (93%), followed by sputum (72%), nasal swab (63%), and pharyngeal swab (32%).<sup>5</sup> False-negative results mainly occurred due to inappropriate timing of sample collection in relation to illness onset and deficiency in sampling technique, especially of nasopharyngeal swabs. Specificity of most of the RT-PCR tests is 100% because the primer design is specific to the genome sequence of SARS-CoV-2. Occasional false-positive results may occur due to technical errors and reagent contamination.

## Detection of Antibodies to SARS-CoV-2

COVID-19 infection can also be detected indirectly by measuring the host immune response to SARS-CoV-2 infection. Serological diagnosis is especially important for patients with mild to moderate illness who may present late, beyond the first 2 weeks of illness onset. Serological diag-

nosis also is becoming an important tool to understand the extent of COVID-19 in the community and to identify individuals who are immune and potentially "protected" from becoming infected.

The most sensitive and earliest serological marker is total antibodies, levels of which begin to increase from the second week of symptom onset.<sup>6</sup> Although IgM and IgG ELISA have been found to be positive even as early as the fourth day after symptom onset, higher levels occur in the second and third week of illness.

For example, IgM and IgG seroconversion occurred in all patients between the third and fourth week of clinical illness onset as measured in 23 patients by To et al<sup>7</sup> and 85 patients by Xiang et al.<sup>8</sup> Thereafter IgM begins to decline and reaches lower levels by week 5 and almost disappears by week 7, whereas IgG persists beyond 7 weeks.<sup>9</sup> In a study of 140 patients, combined sensitivity of PCR and IgM ELISA directed at nucleocapsid (NC) antigen was 98.6% vs 51.9% with a single PCR test. During the first 5.5 days, quantitative PCR had a higher positivity rate than IgM, whereas IgM ELISA had a higher positivity rate after day 5.5 of illness.<sup>10</sup>

ELISA-based IgM and IgG antibody tests have greater than 95% specificity for diagnosis of COVID-19. Testing of paired serum samples with the initial PCR and the second 2 weeks later can further increase diagnostic accuracy. Typically, the majority of antibodies are produced against the most abundant protein of the virus, which is the NC. Therefore, tests that detect antibodies to NC would be the most sensitive. However, the receptor-binding domain of S (RBD-S) protein is the host attachment protein, and antibodies to RBD-S would be more specific and are expected to be neutralizing. Therefore, using one or both antigens for detecting IgG and IgM would result in high sensitivity.<sup>7</sup> Antibodies may, however, have cross-reactivity with SARS-CoV and possibly other coronaviruses.

Rapid point-of-care tests for detection of antibodies have been widely developed and marketed and are of variable quality. Many manufacturers do not reveal the nature of antigens used. These tests are purely qualitative in nature and can only indicate the presence or absence of SARS-CoV-2 antibodies. The presence of neutralizing antibodies can only be confirmed by a plaque reduction neutralization test. However, high titers of IgG antibodies detected by ELISA have been shown to positively correlate with neutralizing antibodies.<sup>7</sup> The long-term persistence and duration of protection conferred by the neutralizing antibodies remains unknown.

## Conclusions

Using available evidence, a clinically useful timeline of diagnostic markers for detection of COVID-19 has been devised (**Figure**). Most of the available data are for adult populations who are not immunocompromised. The time course of PCR positivity and seroconversion may vary in children and other groups, including the large population of asymptomatic individuals who go undiagnosed without active surveillance. Many questions remain, particularly how long potential immunity lasts in individuals, both asymptomatic and symptomatic, who are infected with SARS-CoV-2.

## Article Information

[Back to top](#)

**Corresponding Author:** Sundararaj Stanleyraj Jeremiah, MD, Department of Microbiology and Molecular Biodefense Research, Yokohama City University School of Medicine, 3-9 Fukuura, Kanazawa-ku, Yokohama 236-0004, Japan ([rediffjerry@gmail.com](mailto:rediffjerry@gmail.com)).

**Published Online:** May 6, 2020. doi:[10.1001/jama.2020.8259](https://doi.org/10.1001/jama.2020.8259)

**Conflict of Interest Disclosures:** None reported.

## References

1. Nalla AK, Casto AM, Huang MW, et al. Comparative performance of SARS-CoV-2 detection assays using seven different primer/probe sets and one assay kit. *J Clin Microbiol*. 2020;JCM.00557-20. Published online April 8, 2020. doi:[10.1128/JCM.00557-20](https://doi.org/10.1128/JCM.00557-20)  
[PubMed](#) | [Google Scholar](#)
2. Zheng S, Fan J, Yu F, et al. Viral load dynamics and disease severity in patients infected with SARS-CoV-2 in Zhejiang province, China, January-March 2020: retrospective cohort study. *BMJ*. 2020;369:m1443. Published online April 21, 2020. doi:[10.1136/bmj.m1443](https://doi.org/10.1136/bmj.m1443)  
[PubMed](#) | [Google Scholar](#) | [Crossref](#)

3. Wölfel R, Corman VM, Guggemos W, et al. Virological assessment of hospitalized patients with COVID-2019. *Nature*. 2020. Published online April 1, 2020. doi:10.1038/s41586-020-2196-x  
[PubMed](#) | [Google Scholar](#)
4. CDC. Return-to-work criteria for healthcare workers. Updated April 30, 2020. Accessed May 3, 2020.  
<https://www.cdc.gov/coronavirus/2019-ncov/hcp/return-to-work.html>
5. Wang W, Xu Y, Gao R, et al. Detection of SARS-CoV-2 in different types of clinical specimens. *JAMA*. 2020. Published online March 11, 2020. doi:10.1001/jama.2020.3786  
[Article](#) | [PubMed](#) | [Google Scholar](#)
6. Lou B, Li T, Zheng S, et al. Serology characteristics of SARS-CoV-2 infection since the exposure and post symptoms onset. *medRxiv*. Preprint posted March 27, 2020. doi:10.1101/2020.03.23.20041707
7. To KK-W, Tsang OT-Y, Leung W-S, et al. Temporal profiles of viral load in posterior oropharyngeal saliva samples and serum antibody responses during infection by SARS-CoV-2: an observational cohort study. *Lancet Infect Dis*. 2020;20(5):565-574. doi:10.1016/S1473-3099(20)30196-1  
[PubMed](#) | [Google Scholar](#) | [Crossref](#)
8. Xiang F, Wang X, He X, et al. Antibody detection and dynamic characteristics in patients with COVID-19. *Clin Infect Dis*. 2020;ciaa461. Published online April 19, 2020. doi:10.1093/cid/ciaa461  
[PubMed](#) | [Google Scholar](#)
9. Xiao AT, Gao C, Zhang S. Profile of specific antibodies to SARS-CoV-2: the first report. *J Infect*. 2020;S0163-4453(20)30138-9. Published online March 21, 2020. doi:10.1016/j.jinf.2020.03.012  
[PubMed](#) | [Google Scholar](#)
10. Guo L, Ren L, Yang S, et al. Profiling early humoral response to diagnose novel coronavirus disease (COVID-19). *Clin Infect Dis*. 2020;ciaa310. Published online March 21, 2020. doi:10.1093/cid/ciaa310  
[PubMed](#) | [Google Scholar](#)

Comment

## See More About

Infectious Diseases

Pathology and Laboratory Medicine

Coronavirus (COVID19)



Coronavirus Resource Center

## Trending

### Research

Epidemiologic Features and Clinical Course of Patients Infected With SARS-CoV-2 in Singapore

April 21, 2020

### Opinion

The Relative Value Scale Update Committee—Time for an Update

September 24, 2019

## Opinion

Relative Value Units and the Measurement of Physician Performance

September 24, 2019

Select Your Interests

Advertisement

---

### JOB LISTINGS ON JAMA CAREER CENTER®

#### Intensivist

New York City, New York (US)

#### Intensivist

New York City, New York (US)

#### Physician

Hattiesburg, Mississippi

See more at JAMA Career Center

## Others Also Liked

### Erba Mannheim Garners CE Mark for SARS-CoV-2 Antibody Test

staff reporter, 360Dx, 2020

### Hememics Biotechnologies Wins \$638K BARDA Contract to Develop Bluetooth Coronavirus Test

staff reporter, 360Dx, 2020

### Commercial Labs Step up Coronavirus Test Efforts After FDA Guidance

360Dx, 2020

---

Powered by **TREND MD**

## Trending

Assessment of the Contribution of the

Global Miscalculations—Relative Value

Demand for US Hospital Inpatient and

**Work Relative Value Unit Scale to Differences in Physician Compensation**

JAMA Surgery | *Research* | April 15, 2020

**Units and the Value of a Surgeon's Care**

JAMA Surgery | *Opinion* | April 15, 2020

**Intensive Care Unit Beds for Patients With COVID-19**

JAMA Network Open | *Research* |

May 6, 2020